

MYMOMASK – TB N95

Mở đầu năm 2020, khi Covid 19 xâm nhập vào Việt Nam, lúc đó đường lây nhiễm duy nhất thông qua giọt bắn, lúc đó những chiếc khẩu trang vải kháng khuẩn là lá chắn bảo vệ người dân một cách hiệu quả. Khẩu trang y tế quá đắt do nguyên liệu tăng hàng chục lần, trong nước chưa có nhiều đơn vị đầu tư nghiên cứu sản xuất các sản phẩm chất lượng cao như khẩu trang N95.

Chúng tôi là một trong những người đưa khái niệm khẩu trang theo chuẩn N95 đầu tiên của Việt Nam, và mang những sản phẩm chất lượng cao xuất khẩu sang thị trường Mỹ, châu Âu, khẳng định vị thế người Việt.

Sau hơn một năm, sau 4 lần Việt Nam oằn mình chiến đấu với đại dịch, chủng virus cũng đã biến đổi nhiều lần,... khẩu trang vải, khẩu trang y tế 4 lớp đã bộc lộ nhiều yếu điểm trong việc phòng chống đại dịch, khi mật độ vi khuẩn trong không khí dày đặc, việc đeo một chiếc khẩu trang không đạt chuẩn, nguy cơ lây nhiễm rất cao, nhưng nếu đeo một chiếc khẩu trang N95, chi phí dành cho chiếc khẩu trang cũng không phải nhỏ.

Câu hỏi đặt ra, để tối ưu giá thành cho người sử dụng, cần phải tạo ra chiếc khẩu trang kế thừa các ưu điểm của N95, nhưng phải tái sử dụng được nhiều lần, và như thế MyMoMask ra đời.

Vậy MyMoMask có những điểm nổi trội gì? Trước tiên chúng ta hãy so sánh với những chiếc khẩu trang trước kia để tìm ra sự khác biệt.

Các điểm so sánh	Khẩu trang vải kháng khuẩn	Khẩu trang y tế 4 lớp	Khẩu trang N95	Khẩu trang Mymomask
Số lớp cấu thành	2-3 lớp	4 lớp	5 lớp	6 lớp
Kháng giọt bắn	1 lớp	Không có	1 lớp	2 lớp
Độ kháng khuẩn	60-90%	50-70%	>95%	>99%
Ôm khít khuôn mặt	Ôm khít	Bị hở hai bên	Ôm khít	Ôm khít
Độ dễ thở	Dễ thở	Dễ thở	Trợ lực hô hấp kém	Rất dễ thở
Khả năng lọc bụi mịn	Kém	Kém	Tốt	Rất tốt
Cảm nhận khi đeo	Bị dính vào mũi khi hít thở	Cảm giác không an toàn	Bị hằn vết khẩu trang ở phần nếp mũi	Không bị hằn vì có đệm giảm hằn mũi
Tái sử dụng bằng cách giặt	Không an toàn vì với cách này vi khuẩn bị truyền từ lớp ngoài vào lớp trong.			
Tái sử dụng bằng cách hấp tiệt trùng (100 – 120 độ)	Lớp vải kháng khuẩn sử dụng công nghệ tẩm vào cây vải, do đó bị mất khi hấp ở nhiệt độ cao	Lớp giấy meltblown không chịu được độ ẩm, bị phá hủy kết cấu, giảm khả năng lọc khuẩn.	Lớp giấy meltblown không chịu được độ ẩm, bị phá hủy kết cấu, giảm khả năng lọc khuẩn.	Lớp vi lọc cao phân tử chịu được nhiệt độ và độ ẩm. Không bị thay đổi tính chất khi hấp, sấy ở nhiệt độ cao.


MyMoMask được phát triển dựa trên dòng khẩu trang TB N95 đã được kiểm nghiệm tại Nelson Lab Hoa Kỳ theo chỉ tiêu 42 CFR Part 84 về đánh giá khẩu trang theo chuẩn Niosh 95 của Mỹ, 100% bài Test về độ lọc đều đạt > 99%, khẩu trang MyMoMask có khả năng lọc bụi mịn tới 0.26 μm vượt yêu cầu của chuẩn Niosh 95.

Dưới đây là các chứng chỉ đánh giá chất lượng sản phẩm MyMoMask – TB N95:



CHỨNG CHỈ NIOSH - ĐÁNH GIÁ BỞI NELSON LABS

Nelson Labs là phòng Lab số 1 tại Mỹ, chuyên kiểm tra đánh giá chất lượng các sản phẩm, trong đó có đánh giá tiêu chuẩn Niosh. Khẩu trang MyMoMask – TB N95 đã được kiểm nghiệm chất lượng tại phòng Lab này.



Sodium Chloride (NaCl) Aerosol Test Final Report


Reviewer: Huy Phung
 Bioscience &
 Thien Binh Investment, LLC
 1131 Cowley Ave
 Fountain Valley, CA 92708

Test Article: N95
 Study Number: 1386738-S01
 Study Received Date: 02 Apr 2021
 Testing Facility: Nelson Laboratories, LLC
 6280 S. Rowwood Rd.
 Salt Lake City, UT 84123 U.S.A.
 Test Procedure(s): Standard Test Protocol (STP) Number: STP014 Rev 09
 Deviation(s): None

Summary: This procedure was performed to evaluate particulate filter penetration as specified in 42 CFR Part 84 and TEB-APR-STP-0059 for requirements on a N95 respirator. Respirators were conditioned from tested for particle penetration against a polystyrene, sodium chloride (NaCl) particulate aerosol. The challenge aerosol was dried, neutralized, and passed through the test article at a concentration not exceeding 200 mg/m³. The initial airflow resistance and particle penetration for each respirator was determined.


According to 42 CFR Part 84.84, pretesting must be performed by all applicants as part of the application process with NIOSH. Results seen below are part of that pretesting and must be submitted to and accepted by NIOSH for respirator approval.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.



Curtis Gowen electronically approved
 Study Director
 04 May 2021 22:15 (-05:00)
 Study Completion Date and Time

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com
 PFE01-002 Rev 4
 Page 1 of 3



Sodium Chloride (NaCl) Aerosol Test Final Report

Study Number: 1386738-S01
 Sodium Chloride (NaCl) Aerosol Test Final Report


Reviewer: Huy Phung
 Bioscience &
 Thien Binh Investment, LLC
 1131 Cowley Ave
 Fountain Valley, CA 92708

Results: The NIOSH N95 filter efficiency as stated in 42 CFR Part 84.181 is a minimum efficiency for each filter of 95% (95% penetration). The test articles submitted by the sponsor conform to the NIOSH N95 criteria for filter efficiency.

Test Article Number	Connected* Initial Airflow Resistance (mm H ₂ O)	Maximum Particle Penetration (%)	Filtration Efficiency (%)
1	28.8	0.750	99.210
2	28.2	0.907	99.093
3	28.6	0.702	99.208
4	29.8	0.868	99.134
5	29.2	1.20	99.082
6	29.8	0.853	99.047
7	28.3	1.01	99.204
8	32.1	1.18	99.127
9	32.5	1.28	99.371
10	30.0	1.08	99.00
11	28.2	0.952	99.048
12	29.6	1.03	99.216
13	28.8	1.01	99.121
14	29.3	0.973	99.027
15	29.6	0.953	99.047
16	29.7	0.957	99.043
17	28.4	0.814	99.198
18	28.5	0.828	99.171
19	28.3	0.820	99.080
20	29.1	1.15	99.286

* The final airflow resistance value for each test article was determined by subtracting out the background resistance from the system.

Test Method Acceptance Criteria: The filter tester must pass the "Tester Set Up" procedure. The airflow resistance and particle penetration of the reference material must be within the limits set by the manufacturer.



Adam Brigham electronically approved
 Study Director
 18 Feb 2021 20:21 (-05:00)
 Study Completion Date and Time

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com
 PFE01-002 Rev 4
 Page 2 of 3



Sodium Chloride (NaCl) Aerosol Test Final Report

Study Number: 1386738-S01
 Sodium Chloride (NaCl) Aerosol Test Final Report

Reviewer: Huy Phung
 Bioscience &
 Thien Binh Investment, LLC
 1131 Cowley Ave
 Fountain Valley, CA 92708

Filter Test Procedure: Prior to testing, respirators were taken out of their packaging and placed in an environment of 85 ± 5% relative humidity (RH) and 38 ± 2.5°C for 25 ± 1 hours.

The filter tester used in testing was a TSI[®] CERTIFEST[®] Model 8130 Automated Filter Tester that is capable of efficiency measurements of up to 99.999%. It produces a particle size distribution with a count median diameter of 0.075 ± 0.020 microns (µm) and a geometric standard deviation not exceeding 1.85 µm. The mass median diameter was approximately 0.26 µm, which is generally accepted as the most penetrating aerosol size. The reservoir was filled with a 2% NaCl solution and the instrument allowed a minimum warm-up time of 30 minutes. The main regulator pressure was set to 75 ± 5 pounds per square inch (PSI). The filter holder regulator pressure was set to approximately 35 psi. The NaCl aerosol generator pressure was set to approximately 30 psi and the make-up airflow rate was set to approximately 70 liters per minute (L/min).

The NaCl concentration of the test aerosol was determined in mg/m³ by a gravimetric method prior to the test assessment. An entire respirator was mounted on a test fixture, placed into the filter holder, and the NaCl aerosol passed through the outside surface of the test article at a continuous airflow rate of 85 ± 4 L/min. In accordance with NIOSH policy, three respirators were challenged until 200 ± 5 mg of NaCl had contacted each test article. Based upon the test pattern of NIOSH Type 1, the initial penetration reading of the remaining 17 respirators was recorded.



Adam Brigham electronically approved
 Study Director
 18 Feb 2021 20:21 (-05:00)
 Study Completion Date and Time

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com
 PFE01-002 Rev 4
 Page 3 of 3



Determination of Inhalation and Exhalation Resistance for Air-Purifying Respirators Final Report

Study Number: 1386738-S01
 Sodium Chloride (NaCl) Aerosol Test Final Report

Reviewer: Huy Phung
 Bioscience &
 Thien Binh Investment, LLC
 1131 Cowley Ave
 Fountain Valley, CA 92708

Test Article: N95
 Study Number: 1386738-S01
 Study Received Date: 02 Apr 2021
 Testing Facility: Nelson Laboratories, LLC
 6280 S. Rowwood Rd.
 Salt Lake City, UT 84123 U.S.A.
 Test Procedure(s): Standard Test Protocol (STP) Number: STP0145 Rev 05
 Deviation(s): None

Summary: This procedure was performed to evaluate the differential pressure of non-powered air-purifying particulate respirators in accordance with 42 CFR Part 84.180. The air exchange differential or breathability of respirators was measured for inhalation resistance using NIOSH procedure TEB-APR-STP-0007 and exhalation resistance with NIOSH procedure TEB-APR-STP-0003. The differential pressure technique is a simple application of a basic physical principle employing a water manometer differential upstream and downstream of the test material, at a constant flow rate.

According to 42 CFR Part 84.84, pretesting must be performed by all applicants as part of the application process with NIOSH. Results seen below are part of that pretesting and must be submitted to and accepted by NIOSH for respirator approval.

The inhalation resistance criteria as stated in 42 CFR Part 84.180 is an initial inhalation not exceeding 35 mm water column height pressure. The test articles submitted by the sponsor conform to this NIOSH criterion for airflow resistance.

The exhalation resistance criteria as stated in 42 CFR Part 84.180 is an initial exhalation not exceeding 25 mm water column height pressure. The test articles submitted by the sponsor conform to this NIOSH criterion for airflow resistance.

All test method acceptance criteria were met. Testing was performed in compliance with US FDA good manufacturing practice (GMP) regulations 21 CFR Parts 210, 211 and 820.

Test Article Number	Inhalation Resistance (mm H ₂ O)	Exhalation Resistance (mm H ₂ O)
1	21.4	17.2
2	21.4	17.5
3	22.1	17.7




Adam Brigham electronically approved
 Study Director
 18 Feb 2021 20:21 (-05:00)
 Study Completion Date and Time

801-290-7500 | nelsonlabs.com | sales@nelsonlabs.com
 PFE01-002 Rev 3
 Page 1 of 2

CHỨNG CHỈ FFP2 – ĐÁNH GIÁ BỞI SGS ANH QUỐC

SGS tại Anh Quốc là một trong những phòng Lab hiện đại nhất trong hệ thống SGS toàn cầu. Đây là một trong những phòng Lab uy tín nhất trên thế giới trong việc kiểm tra, đánh giá các bài test theo chuẩn.




TEST REPORT NO: 168968 Date: 26 April 2021

THIEN BINH INVESTMENT JOINT STOCK COMPANY
200 MAI ANH TUAN STREET
THANH CONG WARD
BA DINH DISTRICT
HANOI CITY
VIETNAM

The following sample(s) was/were submitted and identified by/on behalf of the client as:

Retailer: PPE Testing
Description of article: TB FFP2 NR MEDICAL FACE MASKS
Retailer style number: WHITE
Retailer Standard Number: UNKNOOWN
Order No./Buyer: VNHL2103005374HG
Quality/Fibre Composition: UNKNOOWN
Date Sample Received/Test Started: 01 April 2021

Tests	Pass	Fail	Remarks
EN149 PPE Face masks	Pass		See results

Signature: 
Lucy Peberdy Technologist

For and on behalf of
SGS United Kingdom Ltd


All samples are conditioned to ISO 139 where conditioning is required (unless otherwise stated)

SGS does not verify authenticity of any Brand/Trademark of products. Buyers must check if the product is genuine with the Brand/Trademark owner directly.

This document is issued by the Company under its General Conditions of Service available at www.sgs.com/terms_and_conditions.htm, for electronic format documents, subject to terms and conditions for Electronic Documents at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings on the basis of the information only and affords the benefit of a qualified opinion only. The Company's sole responsibility is to check and issue documents based on samples submitted for inspection from the issuing party. The Company is not liable for any consequences arising from the use of the information contained herein. This document cannot be reproduced or stored in a retrieval system without prior written approval of the Company. Any unauthorised alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire LE8 6JH | t+44 (0)116 284 6780 | f+44 (0)116 284 6781 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1189985 | Registered Office: Rosemoor Business Park, Ebbw Vale, Gwent NP23 5SE

Page 1 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No:** 168968

Test Results

***Personal Protective Equipment – Respiratory Protective Devices – Filtering Half Masks to Protect against Particles – Requirements, Testing, Marking**
EN 149:2001+A1:2009

Clause 7.4 Packaging
(EN 149:2001+A1:2009 Clause 8.2)

Test Requirement	Results	Comment
Particle filtering half masks shall be offered for sale packaged in such a way that they are protected against mechanical damage and contamination before use	Comply	Pass

Clause 7.5 Material
(EN 149:2001+A1:2009 Clause 8.2 & 8.3.1 & 8.3.2)

Test Requirement	Results	Comment
Materials used shall be suitable to withstand handling and wear over the period for which the particle filtering half mask is designed to be used	Comply	Pass
After undergoing the conditioning described with 8.3.1 none of the particle filtering half masks shall have suffered mechanical failure of the facepiece or straps	Comply	
When conditioned in accordance with 8.3.1 and 8.3.2 the particle filtering half mask shall not collapse	Comply	
Any material from the filter media released by the air flow through the filter shall not constitute a hazard or nuisance for the wearer	Comply	

Clause 7.6 Cleaning and Disinfecting
(EN 149:2001+A1:2009 Clause 8.4 & 8.5 & 8.11)

Test Requirement	Results	Comment
If the particle filtering half mask is designed to be re-usable, the materials used shall withstand the cleaning and disinfecting agents and procedures to be specified by the manufacturer. With reference to 7.9.2 after cleaning and disinfecting the re-usable particle filtering half mask shall satisfy the penetration requirement of the relevant class	Not applicable (Not designed to be reusable)	N/A


Clause 7.7 Practical Performance
(EN 149:2001+A1:2009 Clause 8.4)

Test Requirement	Results	Comment
The particle filtering half mask shall undergo practical performance tests under realistic conditions. These general tests serve the purpose of checking the equipment for imperfections that cannot be determined by the tests described elsewhere in this standard	No imperfections	Pass

This document is issued by the Company under its General Conditions of Service available at www.sgs.com/terms_and_conditions.htm, for electronic format documents, subject to terms and conditions for Electronic Documents at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings on the basis of the information only and affords the benefit of a qualified opinion only. The Company's sole responsibility is to check and issue documents based on samples submitted for inspection from the issuing party. The Company is not liable for any consequences arising from the use of the information contained herein. This document cannot be reproduced or stored in a retrieval system without prior written approval of the Company. Any unauthorised alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire LE8 6JH | t+44 (0)116 284 6780 | f+44 (0)116 284 6781 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1189985 | Registered Office: Rosemoor Business Park, Ebbw Vale, Gwent NP23 5SE

Page 2 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No:** 168968

Test Results

Clause 7.8 Finish of Parts
(EN 149:2001+A1:2009 Clause 8.2)

Test Requirement	Results	Comment
Parts of the device likely to come into contact with the wearer shall have no sharp edges or burrs	No sharp edges or burrs	Pass

Clause 7.8.4 Total Inward Leakage
(EN 149:2001+A1:2009 Clause 8.5)

Test Requirement	Results	Comment
The total inward leakage consists of three components: face seal leakage, exhalation valve leakage (if exhalation valve fitted) and filter penetration.	Detail refer to Appendix 1	Meet FFP1 Meet FFP2

For particle filtering half masks fitted in accordance with the manufacturer's information at least 40 out of the 50 individual exercise results (i.e. 10 subject x 5 exercises) for total inward leakage shall not be greater than:
25% for FFP1, 11% for FFP2, 5% for FFP3
and, in addition, at least 8 out of 10 individual wearer arithmetic means for the total inward leakage shall be not greater than:
22% for FFP1, 8% for FFP2, 2% for FFP3

Appendix 1: Summarization of Test Data


Inward Leakage Test Data

Subject	Sample No.	Condition	Waik (%)	Head Side(side)(%)	Head up/down(%)	Talk (%)	Waik (%)	Mean (%)
Zhou	1	A.R	6.59	6.19	6.05	6.85	5.83	6.30
Liu	2	A.R	7.22	6.52	7.24	7.67	6.81	7.09
Lu	3	A.R	6.32	6.46	6.50	6.94	5.99	6.44
Wang	4	A.R	5.53	4.63	4.70	5.91	5.05	5.16
Shao	5	A.R	7.14	7.20	7.55	7.84	8.78	7.30
Ding	6	T.C	5.51	5.21	5.45	5.72	5.54	5.49
Li	7	T.C	7.39	7.45	7.65	7.86	7.36	7.54
Chen	8	T.C	5.33	5.09	5.15	6.73	6.04	5.67
Song	9	T.C	6.49	6.24	6.89	7.21	6.59	6.88
Ye	10	T.C	7.79	7.30	7.61	8.06	7.52	7.66

This document is issued by the Company under its General Conditions of Service available at www.sgs.com/terms_and_conditions.htm, for electronic format documents, subject to terms and conditions for Electronic Documents at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings on the basis of the information only and affords the benefit of a qualified opinion only. The Company's sole responsibility is to check and issue documents based on samples submitted for inspection from the issuing party. The Company is not liable for any consequences arising from the use of the information contained herein. This document cannot be reproduced or stored in a retrieval system without prior written approval of the Company. Any unauthorised alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire LE8 6JH | t+44 (0)116 284 6780 | f+44 (0)116 284 6781 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1189985 | Registered Office: Rosemoor Business Park, Ebbw Vale, Gwent NP23 5SE

Page 4 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No:** 168968

Test Results

Facial Dimension (mm)

Subject	Face length	Face width	Face depth	Mouth width
Chen	125	150	120	58
Lu	115	132	107	48
Zhou	115	135	106	52
Li	125	130	107	46
Luo	125	136	100	43
Zheng	128	140	112	55
Wang	120	147	103	48
Song	120	140	100	50
Bao	130	134	104	50
Ding	134	150	110	52
Liu	120	135	117	50
Ye	126	137	106	52

Clause 7.9.2 Penetration of Filter Material
(EN 149:2001+A1:2009 Clause 8.11 & EN 13274-7:2019)

Test Requirement	Results	Comment
The penetration of the filter of the particle filtering half mask shall meet the requirements of the following table	Detail refer to Appendix 2	Meet FFP2


Classification	Maximum penetration of test aerosol	
	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min
FFP1	max. %	max. %
FFP2	20	6
FFP3	1	1

This document is issued by the Company under its General Conditions of Service available at www.sgs.com/terms_and_conditions.htm, for electronic format documents, subject to terms and conditions for Electronic Documents at http://www.sgs.com/terms_and_conditions.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings on the basis of the information only and affords the benefit of a qualified opinion only. The Company's sole responsibility is to check and issue documents based on samples submitted for inspection from the issuing party. The Company is not liable for any consequences arising from the use of the information contained herein. This document cannot be reproduced or stored in a retrieval system without prior written approval of the Company. Any unauthorised alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire LE8 6JH | t+44 (0)116 284 6780 | f+44 (0)116 284 6781 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1189985 | Registered Office: Rosemoor Business Park, Ebbw Vale, Gwent NP23 5SE

Page 4 of 11

CHỨNG CHỈ FFP2 – ĐÁNH GIÁ BỞI SGS ANH QUỐC



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No:** 168968

Test Results

Appendix 2: Summarization of Test Data


Penetration of filter material

Aerosol	Condition	Sample No	Penetration (%)
Sodium chloride test	As received	1	0.575
		2	0.531
		3	0.492
	Simulated wearing treatment	4	0.485
		5	0.563
		6	0.572
Paraffin oil test	As received	7	2.261
		8	2.315
		9	2.016
	Simulated wearing treatment	10	0.662
		11	0.795
		12	0.683
Mechanical strength + Temperature conditioned	As received	13	0.651
		14	0.694
	Simulated wearing treatment	15	0.703
		16	7.975
Mechanical strength + Temperature conditioned	17	7.316	
	18	7.897	

Flow conditioning: Single filter: 95.0 l/min

This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditional to, its electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms and Conditions applicable to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the report only and other the terms of Client's instructions. Any of the Company's responsibility to its Client and this document does not constitute part of a transaction nor warranty of all rights and obligations under the transaction documents. This document cannot be copied, stored or used in any way without the prior approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offences may be prosecuted to the full extent of the law. Unless otherwise stated the results shown in this test report only refer to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whittlesea, Leicestershire, LE18 2RH | t: +44 (0)116 284 6780 | f: +44 (0)116 284 6761 | www.sgs.com
 Member of the SGS Group (SGS SA)
 Registered in England No: 1193366 | Registered Office: Rosemoor Business Park | Ebbw Vale, South Wales | CF23 5JF
 Page 6 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No:** 168968

Test Results

Clause 7.10 Compatibility with Skin
(EN 149:2001+A1:2009 Clause 8.4 & 8.5)

Test Requirement	Results	Comment
Materials that may come into contact with the wearer's skin shall not be known to be likely to cause irritation or any other adverse effect to health	No irritation or any other adverse effect to health	Pass

Clause 7.11 Flammability
(EN 149:2001+A1:2009 Clause 8.6)

Test Requirement	Results	Comment
The material used shall not present a danger for the wearer and shall not be of highly flammable nature	Detail refer to Appendix 3	Pass

When tested, the particle filtering half mask shall not burn or not to continue to burn for more than 5s after removal from the flame

Appendix 3: Summarization of Test Data

Flammability

Condition	Sample No.	Result
As received	1	Nil
Temperature conditioned	2	Nil
	3	Nil
	4	Nil

Clause 7.12 Carbon Dioxide Content of the Inhalation Air
(EN 149:2001+A1:2009 Clause 8.7)

Test Requirement	Results	Comment
The carbon dioxide content of the inhalation air (dead space) shall not exceed an average of 1.0% (by volume)	Detail refer to Appendix 4	Pass

Appendix 4: Summarization of Test Data


Carbon Dioxide Content of the Inhalation Air

Condition	Sample No	Result (%)
As received	1	0.5833
	2	0.5805
	3	0.5819

Mean value: 0.58

This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditional to, its electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms and Conditions applicable to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the report only and other the terms of Client's instructions. Any of the Company's responsibility to its Client and this document does not constitute part of a transaction nor warranty of all rights and obligations under the transaction documents. This document cannot be copied, stored or used in any way without the prior approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offences may be prosecuted to the full extent of the law. Unless otherwise stated the results shown in this test report only refer to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whittlesea, Leicestershire, LE18 2RH | t: +44 (0)116 284 6780 | f: +44 (0)116 284 6761 | www.sgs.com
 Member of the SGS Group (SGS SA)
 Registered in England No: 1193366 | Registered Office: Rosemoor Business Park | Ebbw Vale, South Wales | CF23 5JF
 Page 6 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No:** 168968

Test Results

Clause 7.13 Head Harness
(EN 149:2001+A1:2009 Clause 8.4 & 8.5)

Test Requirement	Results	Comment
The head harness shall be designed so that the particle filtering half mask can be donned and removed easily	Comply	Pass
The head harness shall be adjustable or self-adjusting and shall be sufficiently robust to hold the particle filtering half mask firmly in position and be capable of maintaining total inward leakage requirements for the device	Comply	

Clause 7.14 Field of Vision
(EN 149:2001+A1:2009 Clause 8.4)


Test Requirement	Results	Comment
The field of vision is acceptable if determined so in practical performance tests	Comply	Pass

Clause 7.15 Exhalation Valve(s)
(EN 149:2001+A1:2009 Clause 8.2 & 8.9.1 & 8.3.4 & 8.8)

Test Requirement	Results	Comment
(a) A particle filtering half mask may have one or more exhalation valve(s), which shall function correctly in all orientations	Not applicable due to No exhalation valve	N/A
(b) If an exhalation valve is provided it shall be protected against or be resistant to dirt and mechanical damage and may be shrouded or may include any other device that may be necessary for the particle filtering half mask to comply with 7.9	Not applicable due to No exhalation valve	
(c) Exhalation valve(s), if fitted, shall continue to operate correctly after a continuous exhalation flow of 300 l/min over a period of 30s	Not applicable due to No exhalation valve	
(d) When the exhalation valve housing is attached to the facebuckle, it shall withstand axially a tensile force of 10N applied for 10s	Not applicable due to No exhalation valve	

This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditional to, its electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms and Conditions applicable to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the report only and other the terms of Client's instructions. Any of the Company's responsibility to its Client and this document does not constitute part of a transaction nor warranty of all rights and obligations under the transaction documents. This document cannot be copied, stored or used in any way without the prior approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offences may be prosecuted to the full extent of the law. Unless otherwise stated the results shown in this test report only refer to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whittlesea, Leicestershire, LE18 2RH | t: +44 (0)116 284 6780 | f: +44 (0)116 284 6761 | www.sgs.com
 Member of the SGS Group (SGS SA)
 Registered in England No: 1193366 | Registered Office: Rosemoor Business Park | Ebbw Vale, South Wales | CF23 5JF
 Page 7 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No:** 168968

Test Results

Clause 7.16 Breathing Resistance
(EN 149:2001+A1:2009 Clause 8.9)

Test Requirement	Results	Comment
The penetration of the filter of the particle half mask shall meet the requirements of the following table	Detail refer to Appendix 5	Meet FFP1 Meet FFP2

Classification	Maximum permitted resistance (mbar)					
	Inhalation			Exhalation		
	30 l/min	95 l/min	160 l/min			
FFP1	0.6	2.1	3.0			
FFP2	0.7	2.4	3.0			
FFP3	1.0	3.0	3.0			

Appendix 5: Summarization of Test Data

Breathing resistance (mbar)


As received	Flow rate (l/min)					
	1	2	3	4	5	6
Inhalation	A	B	C	D	E	F
	30	93	03	04	04	03
Exhalation	A	B	C	D	E	F
	95	14	14	13	13	13
Simulated wearing treatment	A	B	C	D	E	F
	160	2.0	2.1	2.1	2.0	2.0
Temperature conditioned	A	B	C	D	E	F
	30	03	04	04	03	03
Exhalation	A	B	C	D	E	F
	95	13	14	14	13	13

A: Facing directly ahead; B: Facing vertically upwards; C: Facing vertically downwards; D: Lying on the left side; E: Lying on the right side

This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditional to, its electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms and Conditions applicable to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the report only and other the terms of Client's instructions. Any of the Company's responsibility to its Client and this document does not constitute part of a transaction nor warranty of all rights and obligations under the transaction documents. This document cannot be copied, stored or used in any way without the prior approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offences may be prosecuted to the full extent of the law. Unless otherwise stated the results shown in this test report only refer to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whittlesea, Leicestershire, LE18 2RH | t: +44 (0)116 284 6780 | f: +44 (0)116 284 6761 | www.sgs.com
 Member of the SGS Group (SGS SA)
 Registered in England No: 1193366 | Registered Office: Rosemoor Business Park | Ebbw Vale, South Wales | CF23 5JF
 Page 8 of 11

CHỨNG CHỈ FFP2 – ĐÁNH GIÁ BỞI SGS ANH QUỐC



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No.:** 168968

Test Results

Clause 7.17 Clogging
(EN 149:2001+A1:2009 Clause 8.9 & 8.10)

Test Requirement	Results	Comment																				
Clause 7.17.2 Breathing resistance Valved particle filtering half masks: FFP1: 4 mbar, FFP2: 5mbar, FFP3: 7mbar at 95L/min continuous flow. The exhalation resistance shall not exceed 3 mbar at 160 L/min continuous flow. Valveless particle filtering half masks: After clogging the inhalation and exhalation resistances shall not exceed: FFP1: 3 mbar, FFP2: 4mbar, FFP3: 5mbar at 95L/min continuous flow.	Optional for single shift device only	N/A																				
Clause 7.17.3 Penetration of filter material All types (valved and valveless) of particle filtering half masks claimed to meet the clogging requirement shall also meet the requirement.	Optional for single shift device only	N/A																				
<table border="1"> <thead> <tr> <th rowspan="2">Classification</th> <th colspan="2">Maximum penetration of test aerosol</th> </tr> <tr> <th>Sodium chloride test 95 l/min</th> <th>Paraffin oil test 95 l/min</th> </tr> </thead> <tbody> <tr> <td></td> <td>%</td> <td>%</td> </tr> <tr> <td></td> <td>Max.</td> <td>Max.</td> </tr> <tr> <td>FFP1</td> <td>20</td> <td>20</td> </tr> <tr> <td>FFP2</td> <td>5</td> <td>5</td> </tr> <tr> <td>FFP3</td> <td>1</td> <td>1</td> </tr> </tbody> </table>	Classification	Maximum penetration of test aerosol		Sodium chloride test 95 l/min	Paraffin oil test 95 l/min		%	%		Max.	Max.	FFP1	20	20	FFP2	5	5	FFP3	1	1		
Classification		Maximum penetration of test aerosol																				
	Sodium chloride test 95 l/min	Paraffin oil test 95 l/min																				
	%	%																				
	Max.	Max.																				
FFP1	20	20																				
FFP2	5	5																				
FFP3	1	1																				


Clause 7.18 Demountable Parts
(EN 149:2001+A1:2009 Clause 8.2)

Test Requirement	Results	Comment
All demountable parts (if fitted) shall be readily connected and secured, where possible by hand	No demountable parts	N/A

*Sub contracted to a Ilac-MRA & CNAS accredited lab



This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditions for use. For electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms/FormsandConditions.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction clauses set forth therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the laboratory only and other the terms of Client's instructions. Any. The Company's sole responsibility is to its Client and this document does not assume any part in a transaction from which any right and obligation under the transaction documents. This document covers the company's activity only and does not constitute an offer of any financial product. The Company's liability is limited to the extent of the contract or appearance of this document is unaltered and otherwise may be processed to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire, Leicestershire, LE8 6JH | t+44 (0)115 284 6780 | f+44 (0)115 284 6791 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1193988 | Registered Office: Rosemoor Business Park | Ebbw Vale | Powys | CF83 3EJ
Page 9 of 11




Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No.:** 168968

Test Results



This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditions for use. For electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms/FormsandConditions.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction clauses set forth therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the laboratory only and other the terms of Client's instructions. Any. The Company's sole responsibility is to its Client and this document does not assume any part in a transaction from which any right and obligation under the transaction documents. This document covers the company's activity only and does not constitute an offer of any financial product. The Company's liability is limited to the extent of the contract or appearance of this document is unaltered and otherwise may be processed to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire, Leicestershire, LE8 6JH | t+44 (0)115 284 6780 | f+44 (0)115 284 6791 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1193988 | Registered Office: Rosemoor Business Park | Ebbw Vale | Powys | CF83 3EJ
Page 10 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No.:** 168968

Test Results


Product information is provided by applicant without verification of authentication of the brand.

The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report



This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditions for use. For electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms/FormsandConditions.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction clauses set forth therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the laboratory only and other the terms of Client's instructions. Any. The Company's sole responsibility is to its Client and this document does not assume any part in a transaction from which any right and obligation under the transaction documents. This document covers the company's activity only and does not constitute an offer of any financial product. The Company's liability is limited to the extent of the contract or appearance of this document is unaltered and otherwise may be processed to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire, Leicestershire, LE8 6JH | t+44 (0)115 284 6780 | f+44 (0)115 284 6791 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1193988 | Registered Office: Rosemoor Business Park | Ebbw Vale | Powys | CF83 3EJ
Page 11 of 11



Customer: THIEN BINH INVESTMENT JOINT STOCK COMPANY **Test Report No.:** 168968

Test Results

Product information is provided by applicant without verification of authentication of the brand.

The statement of conformity in this test report is only based on measured values by the laboratory and does not take their uncertainties into consideration.

End of Report

This document is issued by the Company under its General Conditions of Service accessible at www.sgs.com/terms and, conditions for use. For electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms/FormsandConditions.aspx. Attention is drawn to the limitation of liability, indemnification and jurisdiction clauses set forth therein. Any holder of this document is advised that information contained herein reflects the Company's findings at the time of the laboratory only and other the terms of Client's instructions. Any. The Company's sole responsibility is to its Client and this document does not assume any part in a transaction from which any right and obligation under the transaction documents. This document covers the company's activity only and does not constitute an offer of any financial product. The Company's liability is limited to the extent of the contract or appearance of this document is unaltered and otherwise may be processed to the fullest extent of the law. Unless otherwise stated the results shown in this test report refer only to the samples tested and such samples are retained for 90 days only.

SGS United Kingdom Ltd | White Estate, Cambridge Road, Whetstone, Leicestershire, Leicestershire, LE8 6JH | t+44 (0)115 284 6780 | f+44 (0)115 284 6791 | www.sgs.com
Member of the SGS Group (SGS SA)
Registered in England No: 1193988 | Registered Office: Rosemoor Business Park | Ebbw Vale | Powys | CF83 3EJ
Page 11 of 11

CHỨNG CHỈ FDA

16.31, 10/09/2021 Establishment Registration & Device Listing

FDA

FDA Home³ Medical Devices⁴ Databases⁵
Establishment Registration & Device Listing Back To Search Results
New Search

Proprietary Name: TB 99; TB N95 - 1980; TB N95 - 1987; TB N95 - 1991
Classification Name: MASK, SURGICAL
Product Code: FXK⁶
Device Class: 2
Regulation Number: 878.4040⁷
Medical Specialty: General & Plastic Surgery
Registered Establishment Name: THIEN BINH INVESTMENT JOINT STOCK COMPANY⁸
Registered Establishment Number: 3017032058
Owner/Operator: EconBlue⁹
Owner/Operator Number: 10074551
Establishment Operations: Manufacturer; Repackager/Relabeler

Links on this page:

1. <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
2. <http://www.addthis.com/bookmark.php>
3. <https://www.fda.gov/>
4. <https://www.fda.gov/Medical-Devices>
5. <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
6. </scripts/cdrh/cfdocs/cfPCD/classification.cfm?ID=2906>
7. </scripts/cdrh/cfdocs/cfCFR/CFRsearch.cfm?FR=878.4040>
8. </scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=268994>
9. /scripts/cdrh/cfdocs/cfRL/rl.cfm?start_search=1&OwnerOperatorNumber=10074551

Page Last Updated: 09/06/2021
Note: If you need help accessing information in different file formats, see [Instructions for Downloading Viewers and Players](#).
Language Assistance Available: Español | 繁體中文 | Tiếng Việt | ភាសាខ្មែរ | Tagalog | Русский | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | العربية | English
Accessibility Contact FDA Careers FDA Basics FOIA No FEAR Act Nondiscrimination Website Policies / Privacy

FDA
U.S. Food and Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
Ph. 1-888-INFO-FDA (1-888-463-6332)
Contact FDA


For Government For Press

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=723965&pd=FXK> 1/2

16.31, 10/09/2021 Establishment Registration & Device Listing

Combination Products Advisory Committees Science & Research Regulatory Information Safety Emergency Preparedness International Programs News & Events Training and Continuing Education Inspections/Compliance State & Local Officials Consumers Industry Health Professionals FDA Archive

U.S. Department of Health & Human Services

Links on this page:

1. <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
2. <http://www.addthis.com/bookmark.php>
3. <https://www.fda.gov/>
4. <https://www.fda.gov/Medical-Devices>
5. <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
6. </scripts/cdrh/cfdocs/cfPCD/classification.cfm?ID=2906>
7. </scripts/cdrh/cfdocs/cfCFR/CFRsearch.cfm?FR=878.4040>
8. </scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=268994>
9. /scripts/cdrh/cfdocs/cfRL/rl.cfm?start_search=1&OwnerOperatorNumber=10074551

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=723965&pd=FXK> 2/2

16.32, 10/09/2021 Establishment Registration & Device Listing

FDA

FDA Home³ Medical Devices⁴ Databases⁵
Establishment Registration & Device Listing Back To Search Results
New Search


Proprietary Name: TB N95 - 1980; TB N95 - 1987; TB N95 - 1991
Classification Name: RESPIRATOR, SURGICAL
Product Code: MSH⁶
Device Class: 2
Regulation Number: 878.4040⁷
Medical Specialty: General & Plastic Surgery
Registered Establishment Name: THIEN BINH INVESTMENT JOINT STOCK COMPANY⁸
Registered Establishment Number: 3017032058
Owner/Operator: EconBlue⁹
Owner/Operator Number: 10074551
Establishment Operations: Manufacturer; Repackager/Relabeler

Links on this page:

1. <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
2. <http://www.addthis.com/bookmark.php>
3. <https://www.fda.gov/>
4. <https://www.fda.gov/Medical-Devices>
5. <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
6. </scripts/cdrh/cfdocs/cfPCD/classification.cfm?ID=2924>
7. </scripts/cdrh/cfdocs/cfCFR/CFRsearch.cfm?FR=878.4040>
8. </scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=268994>
9. /scripts/cdrh/cfdocs/cfRL/rl.cfm?start_search=1&OwnerOperatorNumber=10074551

Page Last Updated: 09/06/2021
Note: If you need help accessing information in different file formats, see [Instructions for Downloading Viewers and Players](#).
Language Assistance Available: Español | 繁體中文 | Tiếng Việt | ភាសាខ្មែរ | Tagalog | Русский | العربية | Kreyòl Ayisyen | Français | Polski | Português | Italiano | Deutsch | 日本語 | العربية | English
Accessibility Contact FDA Careers FDA Basics FOIA No FEAR Act Nondiscrimination Website Policies / Privacy

FDA
U.S. Food and Drug Administration
10903 New Hampshire Avenue
Silver Spring, MD 20993
Ph. 1-888-INFO-FDA (1-888-463-6332)
Contact FDA


For Government For Press

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=724003&pd=MSH> 1/2

16.32, 10/09/2021 Establishment Registration & Device Listing

Combination Products Advisory Committees Science & Research Regulatory Information Safety Emergency Preparedness International Programs News & Events Training and Continuing Education Inspections/Compliance State & Local Officials Consumers Industry Health Professionals FDA Archive

U.S. Department of Health & Human Services

Links on this page:

1. <http://www.addthis.com/bookmark.php?u508=true&v=152&username=fdomain>
2. <http://www.addthis.com/bookmark.php>
3. <https://www.fda.gov/>
4. <https://www.fda.gov/Medical-Devices>
5. <https://www.fda.gov/medical-devices/device-advice-comprehensive-regulatory-assistance/medical-device-databases>
6. </scripts/cdrh/cfdocs/cfPCD/classification.cfm?ID=2924>
7. </scripts/cdrh/cfdocs/cfCFR/CFRsearch.cfm?FR=878.4040>
8. </scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=268994>
9. /scripts/cdrh/cfdocs/cfRL/rl.cfm?start_search=1&OwnerOperatorNumber=10074551

<https://www.accessdata.fda.gov/scripts/cdrh/cfdocs/cfRL/rl.cfm?rid=724003&pd=MSH> 2/2



Certificate of Compliance

Application for
Medical Device Directive (MDD) 93/42/EEC &
Personal Protective Equipment (PPE) 2016/425/EU.

This is to certify that the product(s):

Face Mask 3, 4, 5 Layers, Medical Face Mask 3 Layers, Medical Face Mask 4 Layers, Medical Face Mask 5 Layers, FFP1, FFP2, FFP3, Medical Protective Clothing, Non-surgical Isolation Gown, Personal Protection Kit, Medical Protective Gown, Examination Powder-Free Nitrile Glove, Examination Powder-Free Latex Glove & Examination Powder-Free Vinyl Glove.

Manufactured by

THIEN BINH INVESTMENT JOINT STOCK COMPANY

Head Office: No. 200, Mai Anh Tuan Street,

Thanh Cong Ward, Ba Dinh District, Hanoi, Vietnam

Factory 1: Km8+500, Thang Long Highway,

An Khanh Commune, Hoai Duc District, Hanoi, Vietnam

Factory 2: Kien Khe Industrial Park, Kien Khe Town,

Thanh Liem District, Ha Nam Province, Vietnam

Complies with the requirement of the

"Test Standard: EN 14683:2019+AC:2019, EN 149:2001+A1:2009,
EN 14126:2003, EN 455-1/2/3/4, EN 420, EN 374-1/2/3/4/5"

has been assessed & found in accordance with the requirements of

**MDD 93/42/EEC as amended by 2007/47/EC Class I (Non - Sterile) &
PPE 2016/425/EU.** QCC is non-notified certification body, issue this

'compliance certificate' after audit of manufacturer product(s) & technical file(s).

This certificate applies to the tested sample only not for whole production.

It's manufacturer sole responsibility to meet all the necessary conformity assessment activities according to **MDD 93/42/EEC & PPE 2016/425/EU** and related standards before placing them on the market & CE mark on the product(s).

Certificate No. : **CE/020606/0121**
Original Certificate Date : 28 - January - 2021
Issue Date : 28 - January - 2021
Expiry Date : 27 - January - 2022

To check this certificate status visit:
"<http://uasl.uk.com/certifiedorganization.html>"



Authorised Signature

For Quality Control Certification

UK Office: 1929, Chynoweth House,
Trevisson Park, Truro-TR48UN, Cornwall, UK



Quality Control Certification accredited by UASL, UK
This certificate doesn't provide the certified organisation with immunity from its legal obligations.
This certificate remains the property of QC Certification to whom it must be returned on request.

CHỨNG CHỈ TRONG NƯỚC: GIẤY CHỨNG NHẬN LƯU HÀNH TỰ DO

BỘ Y TẾ VIỆT NAM
VIET NAM MINISTRY OF
HEALTH

CỘNG HÒA XÃ HỘI CHỦ NGHĨA VIỆT NAM
Độc lập - Tự do - Hạnh phúc
SOCIALIST REPUBLIC OF VIETNAM
Independence - Freedom - Happiness
Ha Noi, date 12 month 08 year 2020

GIẤY CHỨNG NHẬN LƯU HÀNH TỰ DO
CERTIFICATE OF FREE SALES

- 1. Giấy chứng nhận số/Certificate No:** 20000398CFS/BYT-TB-CT
- 2. Sản phẩm/Product(s):** Khẩu trang y tế (Medical Face Mask)
- 3. Chứng loại/Model:** TB N95
- 4. Công ty sở hữu hợp pháp: CÔNG TY CỔ PHẦN ĐẦU TƯ THIÊN BÌNH**
 - **Product(s) Owner:** THIEN BINH INVESTMENT JOINT STOCK COMPANY
 - **Địa chỉ:** Số 200 phố Mai Anh Tuấn, Phường Thành Công, Quận Ba Đình, Thành phố Hà Nội
 - **Address:** No.200, Mai Anh Tuan Street, Thanh Cong Ward, Ba Dinh District, Ha Noi City, Viet Nam
- 5. Công ty sản xuất: CÔNG TY TNHH SẢN XUẤT CÔNG NGHỆ CAO HỒNG PHÁT**
 - **Manufacturer:** HONG PHAT HIGHT TECH PRODUCTION COMPANY LIMITED
 - **Địa chỉ:** Khu công nghiệp Kiện Khê, Thị trấn Kiện Khê, Huyện Thanh Liêm, Tỉnh Hà Nam
 - **Address:** Kien Khe Industrial Park, Kien Khe Town, Thanh Liem District, Ha Nam Province, Vietnam

Văn bản này là để xác nhận rằng các sản phẩm trên tuân theo các tiêu chuẩn liên quan của Việt Nam hoặc tương đương và được phép bán tại Việt Nam. Việc xuất khẩu sản phẩm không bị hạn chế.

This is to certify that the above product(s) comply with the relevant standards of the S.R. Vietnam or equivalent and are allowed to be sold in Vietnam. The exportation of the product(s) is not restricted.

Giấy chứng nhận này có hiệu lực 03 năm kể từ ngày ký.

This certificate is valid for three years from the date of issuance.

KT. BỘ TRƯỞNG
THỨ TRƯỞNG
FOR MINISTER OF HEALTH
DEPUTY MINISTER OF HEALTH


Nguyễn Trường Sơn
Nguyễn Trường Sơn

CHỨNG CHỈ TRONG NƯỚC: CÔNG BỐ ĐỦ ĐKSX

SỞ Y TẾ HÀ NỘI

CỘNG HÒA XÃ HỘI CHỦ NGHĨA VIỆT NAM
Độc lập - Tự do - Hạnh phúc

Số: 200000182/PCBSX-HN

Thành phố Hà Nội, ngày 14 tháng 08 năm 2020

PHIẾU TIẾP NHẬN

Hồ sơ công bố đủ điều kiện sản xuất trang thiết bị y tế

1. Tên cơ sở công bố: CÔNG TY CỔ PHẦN ĐẦU TƯ THIÊN BÌNH
2. Địa chỉ: Số 200 phố Mai Anh Tuấn, Phường Thành Công, Quận Ba Đình, Thành phố Hà Nội
(Sản xuất tại: CÔNG TY CỔ PHẦN ĐẦU TƯ THIÊN BÌNH; Địa chỉ: Km8+500 đại lộ Thăng Long, Thôn An Thọ, Xã An Khánh, Huyện Hoài Đức, Thành phố Hà Nội)
3. Điện thoại: 0334453714 Fax:
4. Số văn bản đề nghị của cơ sở: 08/2020/THIENBINH Ngày: 10/08/2020
5. Tên trang thiết bị y tế cơ sở công bố sản xuất:
Khẩu trang y tế, Găng tay y tế, Quần áo phòng, chống dịch (Bộ trang phục bảo hộ), Áo choàng phẫu thuật kèm mũ và bao bọc giày, Mũ chụp tóc
6. Thành phần hồ sơ:

1	Văn bản phân công, bổ nhiệm người phụ trách chuyên môn của cơ sở sản xuất	x
2	Bản xác nhận thời gian công tác	x
3	Văn bản, chứng chỉ đào tạo về kỹ thuật thiết bị y tế hoặc quản lý thiết bị y tế của người phụ trách chuyên môn	x
4	Văn bản công bố đủ điều kiện sản xuất	x
5	Bản kê khai nhân sự	x
6	Giấy chứng nhận đạt tiêu chuẩn quản lý chất lượng	x
7	Hồ sơ về thiết bị và quy trình sản xuất, kiểm tra chất lượng phù hợp với yêu cầu của loại trang thiết bị y tế mà cơ sở sản xuất	x
8	Hợp đồng với cơ sở đủ năng lực kiểm tra chất lượng để kiểm tra chất lượng trang thiết bị y tế mà cơ sở sản xuất	x
9	Hệ thống theo dõi quản lý quá trình xuất, nhập, tồn kho, sử dụng nguyên liệu là chất ma túy và tiền chất, quá trình xuất, nhập, tồn kho trang thiết bị y tế có chứa chất ma túy và tiền chất và kho bảo quản	x

NGƯỜI TIẾP NHẬN HỒ SƠ



Nguyễn Minh Hải
Chánh Văn phòng

CHỨNG CHỈ TRONG NƯỚC: HIỆU QUẢ LỌC VI KHUẨN



BỘ Y TẾ
VIỆN KIỂM NGHIỆM AN TOÀN VỆ SINH THỰC PHẨM QUỐC GIA
NATIONAL INSTITUTE FOR FOOD CONTROL (NIFC)

Địa chỉ: Số 65 Phạm Thái Duyệt - Mai Dịch - Cầu Giấy - Hà Nội - Việt Nam
Điện thoại: 84-2432262215 / 84-2432262216 Fax: 84-2439335738 * Website: www.nifc.gov.vn

Số: 16752/PKN-VKNQG

PHIẾU KẾT QUẢ KIỂM NGHIỆM
TEST REPORT

1. Tên mẫu: Khẩu trang y tế TB N95
2. Mã số mẫu: 07205118/DV
3. Mô tả mẫu: Mẫu đựng trong túi nilon (10 chiếc/túi) - Số lượng: 1
NSX - HSD: không có; Không có mẫu lưu
4. Số lượng mẫu: 01 mẫu
5. Thời gian lưu mẫu: Không có
6. Ngày nhận mẫu: 03/07/2020
7. Thời gian thử nghiệm: 03/07/2020 - 11/07/2020
8. Nơi gửi mẫu: CÔNG TY CỔ PHẦN ĐẦU TƯ THIỆN BÌNH
Địa chỉ: 200 Phố Mai Anh Tuấn, Phường Thành Công, Quận Ba Đình
Hà Nội
9. Kết quả thử nghiệm: Chỉ tiêu Vi sinh vật

STT	Tên chỉ tiêu	Đơn vị	Phương pháp thử	Kết quả
9.1*	Hiệu quả lọc vi khuẩn	/chiếc	BS EN 14683:2019	99,91 %

Hà Nội, ngày 11 tháng 7 năm 2020

KT. VIỆN TRƯỞNG



PHÓ VIỆN TRƯỞNG
TS. Trần Cao Sơn

1. Các kết quả thử nghiệm ghi trong phiếu này chỉ có giá trị đối với mẫu thí nghiệm.
2. Không được trích dẫn một phần phiếu kết quả thử nghiệm nếu không có sự đồng ý bằng văn bản của Viện Kiểm nghiệm ATVSTPQG.
3. Thông tin về mẫu và khách hàng được ghi theo yêu cầu của nơi gửi mẫu.
4. Chỉ tiêu đánh dấu * là chỉ tiêu được công nhận VILAS.
5. Không nhận khiếu nại trong trường hợp không có mẫu lưu hoặc hết thời hạn lưu mẫu theo quy định.

CHỨNG CHỈ TRONG NƯỚC: BẢNG PHÂN LOẠI TRANG THIẾT BỊ Y TẾ

Công Ty Cổ Phần Tư Vấn Đầu Tư
và Phát Triển Công Nghệ Hà Nội

CỘNG HÒA XÃ HỘI CHỦ NGHĨA VIỆT NAM
Độc lập - Tự do - Hạnh phúc

Số: 64521CN/190000014/PCBPL-BYT

Hà Nội, ngày 17 tháng 08 năm 2021

BẢN KẾT QUẢ PHÂN LOẠI TRANG THIẾT BỊ Y TẾ

Căn cứ Nghị định số 36/2016/NĐ-CP ngày 15 tháng 5 năm 2016 của Chính phủ về quản lý trang thiết bị y tế;

Căn cứ Nghị định số 169/2018 /NĐ-CP ngày 31 tháng 12 năm 2018 của Chính phủ sửa đổi bổ sung một số điều của Nghị định số 36/2016/NĐ-CP ngày 15 tháng 5 năm 2016 của Chính phủ về quản lý trang thiết bị y tế;

Căn cứ Phiếu tiếp nhận hồ sơ công bố đủ điều kiện phân loại số: 190000014/PCBPL-BYT do Bộ Y tế cấp ngày 04/11/2019;

Căn cứ giấy chứng chỉ hành nghề phân loại của người thực hiện phân loại số: 19000482/BYT-CCHNPL ngày cấp: 13/8/2019;

Theo yêu cầu của CÔNG TY CỔ PHẦN ĐẦU TƯ THIÊN BÌNH có địa chỉ Số 200 phố Mai Anh Tuấn, Phường Thành Công, Quận Ba Đình, Thành phố Hà Nội, Việt Nam, chúng tôi phân loại trang thiết bị y tế như sau: Bản kết quả ở trang bên.

Người thực hiện phân loại

Người đại diện hợp pháp của cơ sở thực hiện phân loại



Nguyễn Văn An

Trang thiết bị y tế không phải là trang thiết bị y tế chẩn đoán in vitro



GIÁM ĐỐC
Nguyễn Thị Hà

Nơi nhận:

- Bộ Y tế;
- Sở Y tế các tỉnh, thành phố;
- Hải quan cửa khẩu;
- Lưu: VT

CHỨNG CHỈ TRONG NƯỚC: BẢNG PHÂN LOẠI TRANG THIẾT BỊ Y TẾ

KẾT QUẢ PHÂN LOẠI TRANG THIẾT BỊ Y TẾ
 Số:64521CN/190000014/PCBPL-BYT ngày 17 tháng 08 năm 2021

TT	Tên trang thiết bị y tế	Model	Mã sản phẩm	Hãng, nước sản xuất	Hãng, nước chủ sở hữu	Mục đích sử dụng theo chỉ định của chủ sở hữu	Căn cứ phân loại mức độ rủi ro	Mức độ được phân loại
1	Khẩu trang y tế thông thường	TB99	TB993PLY; TB994PLY;	Công ty Cổ phần Đầu Tư Thiện Bình, Việt Nam	Công ty Cổ phần Đầu Tư Thiện Bình, Việt Nam	Dùng để ngăn ngừa sự lây nhiễm, chống nhiễm khuẩn, bụi trong môi trường thông thường (dạng phẳng) hoặc môi trường lây nhiễm cao (dạng mồm)	Quy tắc 4, Phần II, Thông tư 39/2016/TT -BYT	Loại A
		TBN95	TBN943PLY; TBN954PLY; TBN955PLY; TBN956PLY;					
		TB1980	TB19803PLY; TB19804PLY; TB19805PLY; TB19806PLY;					
		TB1991	TB19913PLY; TB19914PLY; TB19915PLY; TB19916PLY;					
		TBN99KID						

2	Khẩu trang y tế phòng nhiễm khuẩn	TB99NK	TB99NK3PLY; TB994NKPLY; TB99NKKID	Công ty Cổ phần Đầu Tư Thiện Bình, Việt Nam	Công ty Cổ phần Đầu Tư Thiện Bình, Việt Nam	Dùng để ngăn ngừa sự lây nhiễm, chống nhiễm khuẩn, bụi trong môi trường lây nhiễm cao. Ngăn cản và diệt 99,9% vi khuẩn ngay trên bề mặt khẩu trang.	Quy tắc 4, Phần II, Thông tư 39/2016/TT -BYT	Loại A
		TBN95NK	TBN95NK3PLY; TBN95NK4PLY; TBN95NK5PLY; TBN95NK6PLY;					
		TB1980NK	TB1980NK3PLY; TB1980NK4PLY; TB1980NK5PLY; TB1980NK6PLY;					
		TB1991NK	TB1991NK3PLY; TB1991NK4PLY; TB1991NK5PLY; TB1991NK6PLY;					
		TBN95NKKID						

3	Khẩu trang y tế phòng độc hóa chất	TB99HC	TB99HC3PLY; TB99HC4PLY	Công ty Cổ phần Đầu Tư Thiện Bình, Việt Nam	Công ty Cổ phần Đầu Tư Thiện Bình, Việt Nam	Dùng để ngăn ngừa sự lây nhiễm, chống nhiễm khuẩn, bụi trong môi trường. Tác dụng lọc khí độc và hơi độc, tạo luồng khí sạch sau khi đi qua lớp vi lọc than hoạt tính	Quy tắc 4, Phần II, Thông tư 39/2016/TT -BYT	Loại A
		TBN95	TBN95HC3PLY; TBN95HC4PLY; TBN95HC5PLY; TBN95HC6PLY					
		TB1980	TB1980HC3PLY; TB1980HC4PLY; TB1980HC5PLY; TB1980HC6PLY					
		TB1991	TB1991HC3PLY; TB1991HC4PLY; TB1991HC5PLY; TB1991HC6PLY					

CHỨNG CHỈ TRONG NƯỚC: TIÊU CHUẨN TCVN - 8389

<p>BỘ Y TẾ VIỆN TRANG THIẾT BỊ VÀ CÔNG TRÌNH Y TẾ</p>		<p>CỘNG HÒA XÃ HỘI CHỦ NGHĨA VIỆT NAM Độc lập - Tự do - Hạnh phúc</p>	
<p>GIẤY CHỨNG NHẬN THỬ NGHIỆM Số: <i>M.Đ.Đ</i> /VTTB-DGCL</p>			
<p>Cơ quan yêu cầu: Công ty Cổ phần Đầu tư Thiện Bình</p>			
<p>Địa chỉ: Số 200 Mai Anh Tuấn, Phường Thành Công, Quận Ba Đình, Hà Nội, Việt Nam</p>			
<p>Tên sản phẩm: Khẩu trang y tế</p>			
<p>Model: TB N95</p>			
<p>Đơn vị sản xuất: Công ty TNHH Sản xuất Công nghệ cao Hồng Phát</p>			
<p>Địa chỉ: Cụm CN Kien Khê, Thị trấn Kien Khê, Huyện Thanh Liêm, Tỉnh Hà Nam.</p>			
<p>Xuất xứ: Việt Nam</p>			
<p>Người thử nghiệm: Lê Đức Hà</p>			
<p>Tiêu chuẩn thử nghiệm: Theo tiêu chuẩn TCVN 8389-1:2010</p>			
<p>Phương pháp thử nghiệm: Theo TCVN 8389-1:2010</p>			
<p>Kết luận: Khẩu trang đạt tiêu chuẩn TCVN 8389-1:2010 - Khẩu trang y tế thông thường.</p>			
<p>Hà Nội, ngày 06 tháng 07 năm 2020</p>			
<p>VIỆN TRƯỞNG</p>			
<p><i>(Signature)</i></p>			
<p>TS. Lê Thanh Hải</p>			
<p>Địa chỉ: 40 Phương Mai - Đống Đa - Hà Nội * Điện thoại: (024) 62925544 - 3 8523065 - 3 8521248</p>			

<p>KẾT QUẢ THỬ NGHIỆM (Căn cứ theo tiêu chuẩn TCVN 8389-1:2010)</p>				
<p>Thông số kỹ thuật</p>				
TT	Nội dung	Yêu cầu	Đạt	K.Đạt
1	Kết cấu và vật liệu	Vải không dệt, dạng phẳng, có nếp gấp; có lớp vi lọc, thanh nẹp mũi và dây đeo; không có lỗi ngoại quan	X	
2	Hiệu suất lọc đối với sương dầu (%)	Hiệu suất lọc không nhỏ hơn mức 90%	X	
3	Trở lực hô hấp (mmH ₂ O)	Trở lực hô hấp (DP) không lớn hơn mức 9 mmH ₂ O	X	
4	Giới hạn trường nhìn (%)	Giới hạn trường nhìn không lớn hơn mức 6%	X	
5	Khối lượng (g)	Khối lượng không lớn hơn 10 g	X	
6	Giới hạn cho phép của kim loại nặng			
6.1	Hàm lượng Asen (As)	0,17 mg/kg	X	
6.2	Hàm lượng Chì (Pb)	1,00 mg/kg	X	
6.3	Hàm lượng Thủy Ngân (Hg)	0,12 mg/kg	X	
6.4	Hàm lượng Antimon (Sb)	0,10 mg/kg	X	
6.5	Hàm lượng Cadimi (Cd)	0,10 mg/kg	X	
<p>* Kết quả thử nghiệm này chỉ có giá trị đối với các mẫu thử Công ty Cổ phần Đầu tư Thiện Bình cung cấp cho Viện Trang thiết bị và Công trình Y tế ngày 03/07/2020.</p>				
<p>Hà Nội, ngày 06 tháng 07 năm 2020</p>				
<p>SOÁT XÉT</p>				
<p><i>(Signature)</i></p>				
<p>Nguyễn Văn Hùng</p>				
<p><i>(Signature)</i></p>				
<p>Lê Đức Hà</p>				
<p>Địa chỉ: 40 Phương Mai - Đống Đa - Hà Nội * Điện thoại: (024) 62925544 - 3 8523065 - 3 8521248</p>				

CHỨNG CHỈ ISO 13485: 2016



GIẤY CHỨNG NHẬN CERTIFICATE

Chứng nhận hệ thống quản lý chất lượng cho lĩnh vực trang thiết bị y tế của

This is to certify that Medical devices - Quality Management System of

CÔNG TY CỔ PHẦN ĐẦU TƯ THIÊN BÌNH

THIEN BINH INVESTMENT JOINT STOCK COMPANY

Địa chỉ/ Address

Trụ sở: Số 200 phố Mai Anh Tuấn, Phường Thành Công, Quận Ba Đình, Thành phố Hà Nội, Việt Nam/ Office: No. 200 Mai Anh Tuan Street, Thanh Cong Ward, Ba Dinh District, Hanoi City, Vietnam

Xưởng sản xuất: Km8+500 Đại lộ Thăng Long, Thôn An Thọ, Xã An Khánh, Huyện Hoài Đức, Thành phố Hà Nội, Việt Nam/ Factory: Km8 + 500 Thang Long Avenue, An Tho Village, An Khanh Commune, Hoai Duc District, Hanoi City, Vietnam

Đã được đánh giá và phù hợp với các yêu cầu của tiêu chuẩn

Has been assessed and found to conform with the requirements of the following standard

ISO 13485:2016

Cho lĩnh vực/ For the following activities

Sản xuất và kinh doanh khẩu trang, găng tay, bộ quần áo bảo hộ, áo choàng phẫu thuật kèm mũ và bao bọc giày, mũ chụp tóc

Manufacturing and trading in masks, gloves, protective suits, surgical gowns with hats and shoe covers, and hair caps

Chứng chỉ số/ Certification No: 20.13758-MMS/TTP

Ngày cấp/ Issue date: 10/08/2020

Ngày hết hạn/ Expiry date: 09/08/2023

Đại diện TTP/ On behalf of TTP

Giám đốc/ Director



LÊ HOÀNG NHẬT LINH